

# M153 Buttockmate



SAKAMOTO MODEL  
CORPORATION

User's Manual (131126)

## — I N D E X —

●Introduction／Table of Contents	1
●Purpose／Content	2
●Precautions concerning handling	3
●Usage	4
●Storage method	5

### ! Note

Please open your parcel and check the main unit and accessories of the product immediately upon receipt. If the products are delivered damaged or with parts missing, please contact our company or the distributor.

株式会社坂本モデル



## Introduction

---

Thank you for purchasing our product, “Buttockmate.”

Read this user’s manual carefully in order to use the product for a long time appropriately.

This product is closely checked before factory shipment. However, when this product arrives, open it soon, and check if anything on the main unit and accessories are missing or broken.

## Table of contents

---

Introduction／Table of Contents.....	1
Purpose／Contents.....	2
Precautions concerning handling.....	3
Usage.....	4
Storage method.....	5


## Purpose

Training can be performed for the following items:

1. Judgment of intramuscular injection of buttocks  
("Shibusanbu" method - upper outer quadrant of buttocks)

## Contents

1. Main unit (including the drainage bag)
2. Transparent package
3. Wearable belt
4. Dry-cell battery (two AA alkali ones)
5. Syringe for drainage
6. Drainage bag



## **Precautions concerning handling**

1. When training in injection using this product, use a fresh needle tip.  
Otherwise, the epidermis might be degraded.
2. Use water (tap water) only for training. Do not use any fluids other than water.
3. Pouring water is allowed only when the green lamp is lit.

In the other cases, never pour water.

4. Use four AA alkali battery cells.
5. To perform training using this product, the users must be in pairs.
6. When the model gets dirty, wipe off the dirt using a soft cloth with a neutral detergent.

\* Do not use benzine, thinners, or polishing powders.

7. Pay sufficient attention to handling the product because it is fragile.
8. This product does not support Clark's point and Hochstetter's region.  
(M153-1 "Buttockmate" type 2 supports them.)
9. If anything strange is detected when this model is being used, stop using it and notify the dealer or us as soon as possible.
10. This product is a simulator for users to perform advanced training. Nonetheless, it does not guarantee that a similar effect can be obtained by actual treatment.
11. Note in advance that this product is subject to change without notice to improve it.

## Usage

1. Put the drainage bag on the main unit.
  2. Set the battery cells in the main unit.
  3. Turn on the switch on the main unit.
  4. Put the belt with a face faster on the lumbar region.
  5. Put the belt so that its greater trochanter aligns with that of the buddy wearing the product.
  6. Be sure to use water (tap water) if injecting a fluid.  
Pouring water is allowed only when the green lamp is lit.  
In the other cases, never pour water.
  7. Perform training.
  8. It is judged by the embedded lamp and buzzer whether intramuscular injection has succeeded or not.
    - If training was performed in a proper position including depth, the green lamp will light up.
    - If the injected site was not appropriate, the red lamp will light up, and the buzzer will sound.
    - If that position was correct but the needle was inserted too deep, both the red and green lamps will light up, and the buzzer will sound.
- \* The LED lamp lights up as if floating up beyond the skin.

## **Storage method**

1. Most of the water injected in an appropriate position will be drained after a while.

Water drops left in the injected sites or the drain tube will not leak out but dry soon because the material allows air to pass through.

2. After water is disposed out of the drainage bag, it is able to prompt drainage using the drainage syringe if necessary. After that, store the model with the buttocks facing upward.
3. Do not store the product in a place exposed to direct sunlight or hot and humid air.
4. Be sure to take out the battery cells before storing the model.  
Be careful as the battery cells might corrode and cause a power fault if not removed.

---

**SAKAMOTO MODEL**  
C O R P O R A T I O N

---

Head office

34,Higashitakagicho,Shimogamo  
Sakyo-ku, KYOTO 606-0865, JAPAN

Phone: +81-75-701-1135

Fax: +81-75-722-5638

E-mail: [info@sakamoto-model.com](mailto:info@sakamoto-model.com)

URL: <http://sakamoto-model.co.jp>

Overseas Sales Department

1-9-29,Senbahigashi,Minoh City,Osaka  
562-0035, JAPAN

Phone: +81-72-729-8350

Fax: +81-72-729-8385

E-mail: [trade@sakamoto-model.com](mailto:trade@sakamoto-model.com)